



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/003,011	11/01/2001	Roy K. Greenberg	PA-5270-RFB	3255

7590 09/19/2007
Brinks Hofer Gilson & Lione
P.O. Box 10395
Chicago, IL 60610

EXAMINER

PHILOGENE, PEDRO

ART UNIT	PAPER NUMBER
----------	--------------

3733

MAIL DATE	DELIVERY MODE
-----------	---------------

09/19/2007

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Advisory Action Before the Filing of an Appeal Brief	Application No. 10/003,011	Applicant(s) GREENBERG ET AL.	
	Examiner Pedro Philogene	Art Unit 3733	

--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

THE REPLY FILED 10 September 2007 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE.

1. ☒ The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods:

- a) ☒ The period for reply expires 3 months from the mailing date of the final rejection.
b) ☐ The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.

Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

NOTICE OF APPEAL

2. ☐ The Notice of Appeal was filed on _____. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a).

AMENDMENTS

3. ☐ The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because
(a) ☐ They raise new issues that would require further consideration and/or search (see NOTE below);
(b) ☐ They raise the issue of new matter (see NOTE below);
(c) ☐ They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or
(d) ☐ They present additional claims without canceling a corresponding number of finally rejected claims.

NOTE: _____. (See 37 CFR 1.116 and 41.33(a)).


4. ☐ The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324).
5. ☐ Applicant's reply has overcome the following rejection(s): _____.
6. ☐ Newly proposed or amended claim(s) _____ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).
7. ☐ For purposes of appeal, the proposed amendment(s): a) ☐ will not be entered, or b) ☐ will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.
The status of the claim(s) is (or will be) as follows:
Claim(s) allowed: _____.
Claim(s) objected to: _____.
Claim(s) rejected: 1,3-20,22.
Claim(s) withdrawn from consideration: _____.

AFFIDAVIT OR OTHER EVIDENCE

8. ☐ The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e).
9. ☐ The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing of good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1).
10. ☐ The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached.

REQUEST FOR RECONSIDERATION/OTHER

11. ☒ The request for reconsideration has been considered but does NOT place the application in condition for allowance because:
See Continuation Sheet.
12. ☐ Note the attached Information Disclosure Statement(s): (PTO/SB/08) Paper No(s). _____.
13. ☐ Other: _____.


PEDRO PHILOGENE
PRIMARY EXAMINER

Continuation of 11. does NOT place the application in condition for allowance because: In the Remarks filed 9/10/07, applicant stated "The independent claims denote that the elongate control member is a cannula and as such the grasping portion is attached to the cannula. The references cited by the examiner (to) do not teach or suggest this relationship". The examiner begs to differ. Boyle et al disclose (abstract) a filtering assembly having a self-expanding strut assembly that is attached to a guide wire or control assembly; as claimed by applicant. After the filtering assembly collect the debris, the strut assembly forces the filtering to collapsed and grasp the debris for removal. Furthermore, the applicant stated that the filter is not attached to an elongate control member, rather, the filter merely spins about the guide wire. Again the examiner begs to differ. The strut assembly is attached to the elongate member (18); as disclosed in column 6, lines 22-64, and in column 11, line 67, column 12, lines 1-9. As to the size of the loop compare to the size of the deployment site, it all depends on the size of the deployment site where the device is to be used. As to the material used for the loops, it all depends on the suitability for the intended use. As can be seen in the FIGS. the outersheath of Boyle et al has lubricious outer and inner surfaces. As to the rest of applicant's argument, please, see the examiner's response to the amendment in the final action.


PEDRO PHILOGENE
PRIMARY EXAMINER